



Analytical Report No. R-8-00135
Batch code EUJPTO-00005410
Date 21.12.2018



Eurofins Nihon Soken Japan KK

1-1 Kamiyaji Narikawa
JP-9601108 Fukushima - JAPAN

Analytical Report

Sample code Nr.	257-2018-12000240	Sample reception date:	20.12.2018
Sample described as:	Organic Matcha OA Lot. B8-0126	Analysed between:	20.12.2018 – 21.12.2018

RADIOACTIVITY

	Result Unit	LOD
RL009 JP Radioactivity (Cesium-134/137) (10Bq/kg)	Method: Japan MHLW 2012-03-15, Gamma spectrometry	
Activity in Cesium 134	Not Detected Bq/kg	10
Activity in Cesium 137	Not Detected Bq/kg	10

Gamma ray spectrometry with the Germanium semiconductor detector
Instrument:

- SEIKO EG&G Co., Ltd. SEG-EMS (Germasystem)
- ORTEC, Germanium semiconductor detectetor GEM35-70 (Detector)

The tests identified by the two letters code JP are performed in laboratory Eurofins Nihon Soken (Fukushima).
'ND' or 'Not Detected' means below LOD (Limit of detection)

Keijiro Asai
ASM Manager

*****END OF REPORT*****